UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO. CONFIRMATION NO		
10/664,288	09/17/2003	L. Mercer McKinley	101896-206 (DEP5129) 4546		
	7590 06/14/200° CLENNEN & FISH LL	EXAMINER			
	DE CENTER WEST BOULEVARD	ARAJ, MICHAEL J			
BOSTON, MA		ART UNIT	PAPER NUMBER		
			3733		
			MAIL DATE	DELIVERY MODE	
		06/14/2007	PAPER		

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

		1 4 11 41		A P 4/)		1			
	•	Application	1 NO.	Applicant(s)					
Office Action Summary		10/664,288	3	MCKINLEY, L. ME					
		Examiner		Art Unit		•			
		Michael J.		3733					
The MAILING DATE of this communication appears on the cover sheet with the correspondence address Period for Reply									
A SHO WHIC - Exter after - If NO - Failu Any r	ORTENED STATUTORY PERIOD FOR REPLY CHEVER IS LONGER, FROM THE MAILING DATES OF SIX (6) MONTHS from the mailing date of this communication. SIX (6) MONTHS from the mailing date of this communication. The period for reply is specified above, the maximum statutory period were to reply within the set or extended period for reply will, by statute, reply received by the Office later than three months after the mailing and patent term adjustment. See 37 CFR 1.704(b).	ATE OF THI 36(a). In no ever will apply and will , cause the appli	S COMMUNICATION ot, however, may a reply be tin expire SIX (6) MONTHS from cation to become ABANDONE	N. nely filed the mailing date of this cor D (35 U.S.C. § 133).					
Status				•					
1)🖂	Responsive to communication(s) filed on <u>05 Ap</u>	pril 2007.			,	1			
2a) <u></u>	This action is FINAL . 2b) This action is non-final.								
3)	Since this application is in condition for allowance except for formal matters, prosecution as to the merits is								
	closed in accordance with the practice under Ex parte Quayle, 1935 C.D. 11, 453 O.G. 213.								
Dispositi	on of Claims								
 4) Claim(s) 1-16 is/are pending in the application. 4a) Of the above claim(s) 4,6,11 and 15 is/are withdrawn from consideration. 5) Claim(s) is/are allowed. 6) Claim(s) 1-3,5,7-10,12-14 and 16 is/are rejected. 7) Claim(s) is/are objected to. 8) Claim(s) are subject to restriction and/or election requirement. 									
Applicati	on Papers								
10)⊠	The specification is objected to by the Examine The drawing(s) filed on <u>08 February 2007</u> is/are Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct	e: a) acced are aving (s) be tion is require	e held in abeyance. Sed of the drawing(s) is ob	e 37 CFR 1.85(a). jected to. See 37 CF	R 1.121(d).				
11)	The oath or declaration is objected to by the Ex	kaminer. No	te the attached Office	Action or form PT	O-152.				
Priority u	ınder 35 U.S.C. § 119								
 12) Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f). a) All b) Some * c) None of: 1. Certified copies of the priority documents have been received. 2. Certified copies of the priority documents have been received in Application No. 3. Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)). * See the attached detailed Office action for a list of the certified copies not received. 									
		•							
2) Notice 3) Infor	et(s) ce of References Cited (PTO-892) ce of Draftsperson's Patent Drawing Review (PTO-948) mation Disclosure Statement(s) (PTO-1449 or PTO/SB/08) er No(s)/Mail Date	·	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:)-152)				

Application/Control Number: 10/664,288

Art Unit: 3733

DETAILED ACTION

Continued Examination Under 37 CFR 1.114

A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on April 5, 2007 has been entered.

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless -

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

Claims 1-3, 5, 12, 13 and 16 are rejected under 35 U.S.C. 102(e) as being anticipated by Rosenberg et al. (U.S. Publication No. 2004/0147936).

Rosenberg et al. disclose a spinal rod approximator device (10) comprising an implant-gripping member (14) having a distal portion that extends in a direction substantially transverse to a longitudinal axis (A) of a proximal portion of the implant-gripping member, where the distal portion includes opposing legs (22a, 22b), formed on

Application/Control Number: 10/664,288

Art Unit: 3733

a substantially planar U-shaped member (best seen in Figure 2), that are adapted to be positioned under a distal end of a rod-receiving member of a spinal implant (100); a rod-engaging member (18), comprising of opposed arms (38a, 38b) each having a rod-receiving recess (46a, 46b), slidably coupled to the implant-gripping member along the longitudinal axis of the proximal portion at a position proximal to the implant-gripping member, the rod-engaging member having a distal portion that extends transverse to a proximal portion (see Fig. 2); and a pusher member (16) coupled to at least one of the implant-gripping member and the rod-engaging member such that the pusher member is effective to move at least one of the implant-gripping member and the rod-engaging member with respect to one another. The rod-engaging and implant-gripping members are offset from the sliding axis (as seen in Fig. 1).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.

Claims 7-10 and 14 are rejected under 35 U.S.C. 103(a) as being unpatentable over Rosenberg et al. (U.S. Publication No. 2004/0147936) in view of Hayes et al. (U.S. Patent No. 5,020,519).

Rosenberg et al. disclose the claimed invention except for the specifics of the pusher member. Hayes et al. disclose a pusher member comprising a threaded rod,

Application/Control Number: 10/664,288

Art Unit: 3733

that includes a handle member, extending through a threaded bore formed in a portion of the implant-griping member and wherein the threaded rod includes a distal end mated to a portion of the rod-engaging member. The pusher member is fixedly, but freely rotably coupled to the implant-gripping member and its releasably, threadably mated to the rod-engaging member. It would have been obvious to one skilled in the art at the time the invention was made to have constructed the approximator of Rosenberg et al. with a pusher member having a threaded rod in view of Hayes et al., which is another way to have a slidable pusher member to balance the seating load on the rod (abstract).

Response to Arguments

Applicant's arguments with respect to claims 1 and 12 have been considered but are most in view of the new ground(s) of rejection.

Conclusion

THIS ACTION IS MADE FINAL. Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire THREE MONTHS from the mailing date of this action. In the event a first reply is filed within TWO MONTHS of the mailing date of this final action and the advisory action is not mailed until after the end of the THREE-MONTH shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any

extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Michael J. Araj whose telephone number is 571-272-5963. The examiner can normally be reached on M-F 8am-5pm.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Eduardo Robert can be reached on 571-272-4719. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

MJA

EDUARDO C./ROBERT SUPERVISORY PATENT EXAMINER